NOTES:

- 1. SUBSTRATE: N-BK7
- 2. CENTERING TOLERANCE (AT 587.6nm): BEAM DEVIATION: ≤2 ARCMIN
- 3. COATING (APPLY ACROSS CLEAR APERTURE):


\$1, \$2, \$3, & \$4: R(AVG) < 0.5% @ 720 - 1080nm \$5: MULTILAYER DIELECTRIC TO ACHIEVE: TRANSMISSION > 45% ± 5% POLARIZATION < 6%




ELEMENTS TO BE CEMENTED WITH NORLAND OPTICAL ADHESIVE NOA 61

6. MARK IDENTIFIES BEAMSPLITTED COATED RIGHT ANGLE PRISM







SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

	\$1	\$2		
SHAPE	PLANO	PLANO		1
SURFACE ACCURACY	1/8 WAVE	1/8 WAVE	THIRD ANGLE PROJECTION	⊕ ∈
SURFACE QUALITY	40-20	40-20		1
CLEAR APERTURE	> 90%	> 90%	ALL DIMS IN	mm

			Edmund Optic	S®
THIRD ANGLE _ PROJECTION		TITLE	35mm NIR, Non-Polarizing Cube Beams	plitter
ALL DIMS IN	mm	DWG NO	47572	SHEET 1 OF 1