NOTES:

1. SUBSTRATE:
   ELEMENT A: N-SSK8 618/498
   ELEMENT B: N-SF10 728/284
2. ROHS COMPLIANT
3. CENTERING TOLERANCE (AT 587.6nm):
   BEAM DEVIATION (HALF ANGLE): <1 ARCMIN
4. COATING (APPLY ACROSS COATING APERTURE)

   S1 & S4: VIS 0°
   R(AVG) ≤ 0.4% FROM 425 - 675nm @ 0° AOI

   S2 & S3: NONE

△5 FINE GRIND SURFACE

6. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
7. FOCAL LENGTH (EFL): 120.00mm±1%
   BACK FOCAL LENGTH (BFL): 111.00mm
8. PROTECTIVE BEVEL AS NEEDED
9. DESIGN WAVELENGTH: 587.6nm

△9 ELEMENTS TO BE CEMENTED WITH NORLAND OPTICAL ADHESIVE NOA61







SECTION A-A

***FOR INFORMATION ONLY:***
**DO NOT MANUFACTURE PARTS TO THIS DRAWING**

| | SPECIFICATIONS AFTER CEMENTING | | | |
|---|---|---|---|---|
| ELEMENT TITLE | ELEMENT A | | ELEMENT B | |
| SURFACE | S1 | S2 | S3 | S4 |
| SHAPE | CONVEX | CONVEX | CONCAVE | CONVEX |
| RADIUS | 65.22 | 62.03 | 62.03 | 1240.67 |
| SURFACE QUALITY | 40 - 20 | 40 - 20 | 40 - 20 | 40 - 20 |
| MIN CLEAR APERTURE | ∅39.00 | ∅39.00 | ∅39.00 | ∅39.00 |
| MIN COATING APERTURE | ∅39.00 | N/A | N/A | ∅39.00 |
| POWER AT 632.8nm | 3.00 RINGS | 3.00 RINGS | 3.00 RINGS | 3.00 RINGS |
| IRREGULARITY AT 632.8nm | 0.50 RINGS | 0.50 RINGS | 0.50 RINGS | 0.50 RINGS |

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE
DIMENSIONS ARE FOR REFERENCE ONLY

Edmund Optics®

| | | | |
|---|---|---|---|
| THIRD ANGLE PROJECTION | | TITLE | 40mm Dia. x 120mm FL, VIS 0 Coated, Achromatic Lens |
| ALL DIMS IN | mm | DWG NO | 47739 |
| | | | SHEET 1 OF 1 |