NOTES:

1. SUBSTRATE

FUSED SILICA

2. SURFACE S2 TO BE PARALLEL WITH S1 WITHIN < 3 ARCMIN


3. COATING (APPLY ACROSS CLEAR APERTURE) \$1: Laser Mirror (355nm) R(ABS) >99.5% @ 355nm @ 45° AOI R(AVG) >99.5% @ 351 - 358nm @ 45° AOI


DAMAGE THRESHOLD, PULSED: 6 J/cm² @ 355nm, 20ns, 20Hz CW: 1MW/cm² @ 355nm S2: NONE



5. POWER, IRREGULARITY, SURFACE QUALITY, AND COATING SPECIFICATIONS APPLY ACROSS CLEAR APERTURE

6. ROHS COMPLIANT





FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

SHAPE	S1 PLANO	S2 PLANO				Edmund Optic	S [®]
SURFACE FLATNESS	λ/10	N/A				Ø25mm, 355nm, 45°, ND:YAG LASER LIN	
SURFACE QUALITY	10-5	FINE GROUND			TITLE	MIRROR	
CLEAR APERTURE	Ø22.5	N/A		I			CUEFT
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN	mm	DWG NO	47324	Sheet 1 Of 1