## NOTES:

- 1. SUBSTRATE FUSED SILICA
- 2. SURFACE S2 TO BE PARALLEL TO SURFACE S1 TO WITHIN <3 ARCMIN
- 3. COATING (APPLY ACROSS COATING APERTURE):


S1: 343 HR Coating R (ABS) > 99.80% @ 343nm @ 0-45° AOI

DAMAGE THRESHOLD, PULSED: 5 J/cm<sup>2</sup>, 20ns, 20Hz @ 343nm CW: 1MW/cm<sup>2</sup> @ 343nm

S2: NONE

4 FINE GRIND SURFACE

- 5. POWER, IRREGULARITY, AND SURFACE QUALITY SPECIFICATIONS APPLY ACROSS CLEAR APERTURE
- 6. CLEAR APERTURE AND COATING APERTURE ARE CENTERED ON SURFACES
- APPLY ARROW ON EDGE WITH LASER ETCH, PENCIL, OR PERMANENT INK POINTING TOWARDS SURFACE \$1



## *FOR INFORMATION ONLY:* DO NOT MANUFACTURE PARTS TO THIS DRAWING

				BINE		
	S1	\$2	]		۹	
SHAPE	PLANO	PLANO			Edmund Optic	<b>C</b> ®
SURFACE QUALITY	10-5	COMMERCIAL POLISH				5
SURFACE FLATNESS	0.10 WAVE	N/A			Ø6.35mm x 4mm,343nm, Yb:YAG MIRR	
MIN CLEAR APERTURE	Ø5.40	N/A		TITLE	45° AOI	
MIN COATING APERTURE	Ø5.72	N/A				CULET
BEVEL	PROTECTIVE AS NEEDED	PROTECTIVE AS NEEDED	ALL DIMS IN mm	DWG NO	38850	Sheet 1 Of 1

## SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

′4.`