
NOTES:


- 1. SUBSTRATE: N-BK7
- 2. S2 TO BE PARALLEL WITH S1 TO <1 ARCMIN
- 3. COATING (APPLY ACROSS CLEAR APERTURE)

\$1: NIR I (600-1050nm) \$2: NIR I (600-1050nm) Ravg ≤0.5% @ 600 - 1050nm



- 5. ROHS COMPLIANT
- 6. TRANSMITTED WAVEFRONT (P-V): $\lambda/4$ (per inch over clear aperture)





FOR INFORMATION ONLY: DO NOT MANUFACTURE PARTS TO THIS DRAWING

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT NOTICE DIMENSIONS ARE FOR REFERENCE ONLY

| | \$1 | S2 | | |
|-----------------|---------------------|---------------------|--|--|
| SHAPE | PLANO PLANO | | | |
| SURFACE QUALITY | 60-40 | 60-40 | | |
| CLEAR APERTURE | 45.00 x 45.00 | 45.00 x 45.00 | | |
| BEVEL | PROTECTED AS NEEDED | PROTECTED AS NEEDED | | |

| | | | Edmund Optics® | | | |
|---|------------------------|------------------------|----------------|---|-------------------|--|
| | THIRD ANGLE PROJECTION | $\phi \Leftrightarrow$ | TITLE | 50mm Sq., 2mm Thick, NIR I Coated λ/4 N-BK7 Window | | |
| , | ALL DIMS IN | mm | DWG NO | 23441 | SHEET 31 OF 48 | |