

SHAPE

BEVEL

SURFACE ACCURACY

SURFACE QUALITY

CLEAR APERTURE

1. SUBSTRATE: FUSED SILICA

2. RoHS: COMPLIANT

3. SURFACE S2 TO BE PARALLEL TO SURFACE S1 TO WITHIN <5 ARCMIN

4. COATING (APPLY ACROSS CLEAR APERTURE) S1 & S2: Ravg < 0.5% @ 350 - 700nm

## 5. FINE GRIND SURFACE

